



(Registered Copy)
TEST REPORT

7-6, Gomak-Ri, Wolgot-Myeon, Gimpo-Si, Gyeonggi-Do, KOREA

TEL 82-31-999-3000 FAX 82-31-999-3001

Report No : TBH-001084

Receipt Date : May.31.2013

Client : Kim young soo

Test Completion Date : Jul.02.2013

Young lim co.

489-4 Gojan-dong, Namdong-gu, Incheon, Korea

Sample : Medical needle safe disposer(YLC-1002)

TEST RESULTS

TEST ITEM	UNIT	SAMPLE	RESULT	TEST METHOD
Bactericidal test(E. coli : Initial)	CFU/Carrier	-	2.8×10 ⁵	By The Client
Bactericidal test(E. coli : After 3min)	CFU/Carrier	-	1.4×10 ² (99.9 %)	By The Client
Bactericidal test(S. aureus : Initial)	CFU/Carrier	-	3.4×10 ⁵	By The Client
Bactericidal test(S. aureus : After 3min)	CFU/Carrier	-	40 (99.9 %)	By The Client

- Test strain : Escherichia coli ATCC 25922

Staphylococcus aureus ATCC 6538

- Decrease rate(%) = [(A-B)/A] × 100

where, A : Plate count of initial time

B : Plate count of designate time

Attachment : Test Report

USAGE : QUALITY CONTROL

- NOTE :
1. The test results on this test report are only limited to the samples and sample names provided by the customer and KTR do not guarantee the quality of all products of the customer.
 2. This test report shall not be used for public relation, advertisement, lawsuit and any other purposes outside the scope of its defined usage.
 3. This test report includes the test result performed in accordance with the test method which is presented by client.

Kim Ok-yeon

Prepared by Kim Ok-yeon

Tel: +82-31-999-3186

E-mail: yeon475@ktr.or.kr

Kim Yoohyun

Reviewed by Kim Yoohyun

Technical Manager

E-mail: yhk002@ktr.or.kr

Registered copy date: Jul.03.2013

Jul.02.2013

Korea Testing & Research Institute

President *Kim Young-soo*

